

VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

JAN 03 2010

		DataMaster cdm S/N		130569
Supervisor/Operator Performing the Verification Procedure:				
A	Name	Mac Burrell	ID #	4756
			Date	12/23/2009
	Agency	Cordova Police Department	Phone #	424-6100
Instrument Location		Cordova Police Department		
B	Alco S/N	X301686	Target Value	0.086
			High Pressure	1000psi
	Alco Test Value Average	0.087	0.087	
	1st Alco	2nd Alco		
Signature		Mac Burrell		CAB 1/25/10
(OVER)				


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (Do Not write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 28th day of January, 2010.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

Of Data Master cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statedwide Breath Alcohol Program

DataMaster cdm S/N

130569

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

Breath Test Record

Serial Number 130569 Model DataMaster
Location CDV1 Cordova Police Dept.

Subject Information

Last Name Verification Of Cal O.L. # 122309
First Name M

Operator Information

Operator Name Burrell/M Case
Operator # 4756 Department CDV1

Test Results

Alco Target .086 Alco Serial X301686
Nominal 0.086 Barometer 30.09 Test Type V
Dry Gas 1 0.087 Time: 20:17:35 Test Date 12/23/2009
Subject Alcohol 0.000 Time: 20:19:06 Test Time 20:16:12
Dry Gas 2 0.087 Time: 20:20:27 Report Y

DIAGNOSTIC CHECK

Diagnostic Test Record

Serial Number 130569 Model DataMaster
Location CDV1 Cordova Police Dept.

Diagnostic Test Results

Test Date 12/23/2009 Computer Okay
Test Time 20:11:16 Program Okay
Software Date 02/20/2001
Sample Chamber (C) 49
Breath Tube (C) 40
Barometer (in) 30.09
Flow Detector Okay
High Speed Pump Okay
Detector Okay
Filter Okay
Quartz Okay
Calibration Okay

I, Nita J.
(1) I am
(2) The
(3) I am
(4) In th
responsi
maintain
(5) The
regular
verification.

OWS:
ime I
withir
n.
hol pi
y trair
calibra
d by a

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23 day of December, 2010.

Carolyn M. Nola
Notary Public, State of Alaska
Commission Expires with Office

BREATH ANALYSIS

.086 ADJUSTED FOR 30.09 in
ALCO TARGET .086 20:16
ALCO TEST .087 20:17
INTERNAL STANDARD VERIFIED 20:17
ALCO TV 30.11 in .087 20:17
BLANK TEST .000 20:18
SUBJECT SAMPLE .000 20:19
ALCO TEST .087 20:20
ALCO TV 30.11 in .087 20:20
BLANK TEST .000 20:21
(Notary Seal Stamp)

HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!##\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNPQRSTUVWXYZ[\]^_`abcde'fghijklmnop
qrstuvwxyz{|}~*